

Product	Medical Power Supply (for building-in)
Applicant	EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez Andheri (E), Mumbai - 400096 India
Manufacturer	EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez Andheri (E), Mumbai - 400096 India
Factory	EOS Power India PVT Ltd. Unit #57, SDF II, Seepz-Sez Andheri (E), Mumbai - 400096 India <input type="checkbox"/> See next page(s)
Ratings	Input: 6.0A, 100-240Vac, 47-63Hz, Output: 350W Max (See Test Report for output rating and variable definition)
Trade mark	
Model / Type Ref.	LFMWLP350--ZXXX-YYY-W-II
Principal characteristics	Cl. II <input type="checkbox"/> See next page(s)
A sample of the product was tested and found to be in conformity with	MED EN 60601-1:2006;A1
Validity	This certificate documents conformity with the standards shown, and also applies as license for use of Nemkos name and certification mark. The certificate and license is valid as long as the applicable conditions are complied with, and provided that any changes to the product are notified to Nemko for acceptance prior to implementation. New standards or amendments to the standards may imply that the product design must be updated and/or that re-testing and re-certification is necessary.
Additional information	<input type="checkbox"/> See next page(s) The abovementioned certified equipment complies with current regulatory requirements regarding electrical safety in Norway and other EU/EEA member states, as far as this can be checked. Compliance with requirements regarding building-in, protection against electric shock and Electromagnetic Compatibility (EMC) must be checked when the equipment is built-in a completed product or forms a part of a complete system.
Additional model(s)	<input checked="" type="checkbox"/> See next page(s)

Date of issue 24-02-2015

Nastaran Vadoodi

Nastaran Vadoodi


Certification Department

Nemko AS

Gaustadalléen 30, P.O. Box 73 Blindern, 0314 Oslo, Norway

TEL +47 22 96 03 30 FAX +47 22 96 05 50 EMAIL info@nemko.com

ENTERPRISE NUMBER NO974404532

Product	Medical Power Supply (for building-in)
Pos. No	1
Model / Type Ref.	MBC350-ZXXXL-YYY-W-2
Trade mark (if different from page 1)	 <p>bel POWER SOLUTIONS & PROTECTION a bel group</p>
Rating	Input: 6.0A, 100-240Vac, 47-63Hz Output: 350W Max (See Test Report for output rating and variable definition)
Principal characteristics	Cl. II

Date of issue 24-02-2015



Nastaran Vadoodi

Certification Department

Nemko AS

Gaustadalléen 30, P.O. Box 73 Blindern, 0314 Oslo, Norway
TEL +47 22 96 03 30 FAX +47 22 96 05 50 EMAIL info@nemko.com
ENTERPRISE NUMBER NO974404532