

Product	Switch mode power supply (for building-in)
Applicant	EOS Power India PVT.Ltd. Unit #57, SDF II, Seepz-Sez, Andheri (East) Mumbai 400 096 India
Manufacturer	EOS Power India PVT.Ltd. Unit #57, SDF II, Seepz-Sez, Andheri (East) Mumbai 400 096 India
Factory	EOS Power India PVT.Ltd. Unit #57, SDF II, Seepz-Sez, Andheri (East) Mumbai 400 096 India <input type="checkbox"/> See next page(s)
Ratings	AC Input: 3 A max. 100 - 240 V~ 47-63 Hz,
Trade mark	
Model / Type Ref.	LFWLP225-ZXXX-YYY
Principal characteristics	Cl. I. Output: 225W max, Refer to Test Report. In model name, Z = 1 denoting the number of main output, XXX = any numeric from 0-9 (representing the output voltage). Y = any alpha or numeric character or blank (denotes minor output variation and /or minor SELV Circuit variations) <input type="checkbox"/> See next page(s)
A sample of the product was tested and found to be in conformity with	OFF EN 60950-1:2006;A11;A1;A12;A2
Validity	This certificate documents conformity with the standards shown, and also applies as license for use of Nemkos name and certification mark. The certificate and license is valid as long as the applicable conditions are complied with, and provided that any changes to the product are notified to Nemko for acceptance prior to implementation. New standards or amendments to the standards may imply that the product design must be updated and/or that re-testing and re-certification is necessary.
Additional information	<input type="checkbox"/> See next page(s) The abovementioned certified equipment complies with current regulatory requirements regarding electrical safety in Norway and other EU/EEA member states, as far as this can be checked. Compliance with requirements regarding building-in, protection against electric shock and

Date of issue 17-11-2014



Juan Z. Kleppenes
Certification Department

Nemko AS
Gaustadalléen 30, P.O. Box 73 Blindern, 0314 Oslo, Norway
TEL +47 22 96 03 30 FAX +47 22 96 05 50 EMAIL info@nemko.com
ENTERPRISE NUMBER NO974404532

Electromagnetic Compatibility (EMC) must be checked when the equipment is built-in a completed product or forms a part of a complete system.

Additional model(s)

See next page(s)

Date of issue 17-11-2014



Juan Z. Kleppenes

Certification Department

Nemko AS

Gaustadalléen 30, P.O. Box 73 Blindern, 0314 Oslo, Norway
TEL +47 22 96 03 30 FAX +47 22 96 05 50 EMAIL info@nemko.com
ENTERPRISE NUMBER NO974404532